

WHAT IS CLAIMED IS:

1. A calibration device for confirming or calibrating a biopolymeric array optical scanner, said device comprising:

a polymer layer comprising at least one fluorescent agent, wherein said device has minimal local and global nonuniformities.

2. The device according to claim 1, wherein said at least one fluorescent agent is distributed substantially uniformly throughout said polymer.

3. The calibration device according to claim 1, wherein said polymer is selected from the group consisting of acrylates, epoxides, urethanes, polycarbonates, polyolefins, polyetherketones, polyesters, polystyrenes, polyethylstyrene, polysiloxanes, and copolymers thereof.

4. The calibration device according to claim 1, wherein said polymer is polymethyl-methacrylate.

5. The calibration device according to claim 1, wherein the thickness of said polymer layer ranges from about 0.25 micron to about 10 microns.

6. The calibration device according to claim 1, wherein the thickness of said polymer layer ranges from about 0.4 micron to about 1 micron.

7. The calibration device according to claim 1, wherein said device comprises a single polymer layer.

8. The calibration device according to claim 1, wherein said device comprises a plurality of polymer layers.

9. The calibration device according to claim 1, wherein said at least one fluorescent agent is present in said polymer in a final concentration ranging from about 1 ppm to about 5000 ppm.

10. The calibration device according to claim 1, wherein said at least one fluorescent agent absorbs and emits light in the portion of the electromagnetic spectrum to which a photomultiplier tube of said optical scanner is sensitive.

11. The calibration device according to claim 1, wherein said at least one fluorescent agent absorbs and emits light in the wavelength range selected from the group consisting of ultraviolet, visible and infrared.

12. The calibration device according to claim 1, wherein said global nonuniformity of said calibration device is less than about 5%.

13. The calibration device according to claim 1, wherein said local nonuniformity of said calibration device is less than about 5%.

14. The calibration device according to claim 1, further comprising at least one region absent said at least one fluorescent agent.

15. The calibration device according to claim 14, wherein said at least one region is photobleached.

16. The calibration device according to claim 1, wherein said polymer layer comprises at least two fluorescent agents.

17. The calibration device according to claim 1, wherein said polymer layer is selected from the group consisting of a spin-coated polymer layer, a draw coated polymer layer, a roller coated polymer layer, an electrodeposited polymer layer and a sprayed polymer layer.

18. A calibration device for confirming or calibrating a biopolymeric array optical scanner, said device comprising:

- (a) a polymer layer comprising at least one fluorescent agent; and
- (b) at least one region in said polymer layer absent said at least one fluorescent agent.









49. The method according to claim 43, wherein said step of calibrating comprises calibrating the focus position of said system.

50. The method according to claim 49, wherein said focus position calibration comprises adjusting the distance between a scanning stage and a lens of said system.

51. The method according to claim 43, wherein said step of calibrating comprises calibrating the dynamic focus of said system.

52. The method according to claim 51, wherein said dynamic focus calibration comprises adjusting the rate of speed at which an optical stage of said system travels.

53. The method according to claim 43, wherein said step of calibrating comprises determining the amount of oscillation in an intensity image and adjusting the rate of speed of said optical stage according to said oscillation data.

54. The method according to claim 43, wherein said step of calibrating comprises calibrating at least one scanner mirror of said system.

55. The method according to claim 54, wherein said at least one scanner mirror calibration comprises adjusting said at least one scanner mirror to synchronize the light beams of said system.

56. The method according to claim 43, further comprising the steps of subtracting a background signal from said obtained fluorescent data to obtain a background corrected value.

57. The method according to claim 43, wherein said fluorescent agent(s) is distributed substantially uniformly throughout said surface.

58. The method according to claim 43, further comprising the step of verifying the jitter of said optical scanning system.

40005598-120404

59. A method for performing a hybridization assay, said method comprising:  
(a) calibrating an optical scanner with a calibration device, wherein said calibration device is selected from the group consisting of a calibration device according to claim 1 and a calibration device according to claim 18,  
(b) performing a hybridization assay with at least one array, and  
(c) scanning said array with said calibrated optical scanner.

60. The method according to claim 59, wherein said calibration device further comprises at least one photobleached region, and said method further comprises verifying the jitter of said optical scanner based on a scan of said photobleached area.

61. A method comprising forwarding data representing a result of a scan obtained by the method of claim 59.

62. The method according to claim 61, wherein said data is transmitted to a remote location.

63. A method comprising receiving data representing a result of an interrogation obtained by the method of claim 61.

64. A method for manufacturing a calibration device, said method comprising spin-coating a composition onto a substrate to produce a calibration device selected from the group consisting a calibration device according to claim 1 and a calibration device according to claim 18.

65. The method according to claim 61, further comprising photobleaching at least one region of said device.

66. A kit for calibrating a biopolymeric array optical scanner, said kit comprising:

(a) at least one device selected from the group consisting of a device according to claim 1 and claim 18; and



(b) a substrate comprising instruction for using said device to calibrate a biopolymeric array optical scanner.

67. A kit for calibrating a biopolymeric array optical scanner, said kit comprising:

(a) at least one device selected from the group consisting of devices according to claim 1 and claim 18; and

(b) an array.

1000590 120401  
104021 0650001